

MINUTES OF THE INFORMATION MEETING

Contract title: Supply of Emergency Medical Services Equipment for HCIVs/General Hospitals and Yumbe Training and simulation centre.

Reference: UGA20003-10015

The information meeting was held on 12th April 2023 at 10:00 AM at the Representation office.

N°	Issue / question raised	Response
1.	We observe that the budget for this procurement was not indicated	Following the recent developments, budgets shall no longer be indicated budgets in the tender document as it was before as this will help us a contracting authority to be more effective and efficient in the procurement process.
2.	Should the financial proposal for the two lots be separated?	Yes, the financial proposals for Lot one should be separate from the financial proposal in Lot two just like it's indicated in the tender document
3.	Power supply for 75 litres autoclave: 380 - 415VAC/50HZ	Power supply should for the autoclave be revised to AC 220 V, 50 Hz
4.	Whether 50 litre capacity oxygen cylinders should be filled with oxygen or not	The delivery should be filled with oxygen
5.	Whether Oxygen purity of 93%±3 should be considered in this order	This can be considered since it will fall within the specified range if 95%±5 is used
6.	Sizes of sutures for chest drainage system that should be considered	Sizes should include nylon 2-0 or 1 and PGA 2-0
7.	The weight of the single patient monitor should be about max. 2.75kg	The weight of the single patient monitor should be about max. 2.75 to 8kg

Done in Enabel Kololo Office, on 12/3/2023