

MINUTES OF THE NON-COMPULSORY INFORMATION MEETING

Contract title: Public contract for the supply, delivery installation, testing, and user training of medical equipment for health facilities

Reference: UGA180371T-10028

The non- compulsory information meeting was held on 16th-June-2023 at 11:00 AM to 12:00 PM at Plot B Kololo.

N°	Issue / question raised	Response
1.	Neonatal Incubator with skin temperature monitor, Spo2 monitor, apnoea detector and drip stand Should the UPS be a stand alone? Is the Neonatal incubator for only babies?	UPS shall be a stand alone
2.	Oxygen Concentrator with double outlet How many litres should we supply?	10 litres
3.	Pulse Oximeter (Neonatal) Is this for only neonatal?	No, it shall be multipurpose not only for the babies
4.	Should we consider the pictures or the specifications	Supply shall depend on the specifications not pictures or a particular brand. The supply shall meet the specifications.