

CLARIFICATIONS

Contract title: Public supply contract for Supply, delivery, and installation of medical equipment items and sundries for KIDA hospital in Kabarole district and Kagando hospital in Kasese district

Reference: UGA21003-10366

Please note that the contracting authority will not further respond to any other question as provided for under point 3.2 "Requesting clarifications". As a reminder, tenders must be received **before 28th Aug 2026 at 10:00 am, Kampala time**. It must be sent to: **Late bids will not be accepted**.

N°	Issue / question raised	Response
1.	Are the offered medical devices required to be already registered/approved by the Uganda National Drug Authority (NDA) at the time of tender submission? If registration is required, can registration/import authorization be completed after contract award and before importation?	If the brands of the equipment are already existing in the local market in Uganda, the tenderer may not register them, however if the brands of the proposed equipment are non-existent in the market, registration may be required by the Uganda National Drug Authority. Yes, registration/import authorisation can be completed after contract award.
2.	Is a Manufacturer Authorization Letter required for products offered by a distributor or international supplier?	Yes, a Manufacturer Authorisation Letter is required. The tenderer shall provide a MAF for the following equipment: 1. Cart-based digital colour doppler ultrasound system 2. Fully automated clinical chemistry analyser 3. Portable point of care haemoglobin analyser 4. Digital portable colour doppler ultrasound system 5. Automated 5-part differential hematology analyser
3.	Would CE certification/FDA approval together with the required ISO/IEC documentation be considered sufficient regulatory evidence at tender stage?	Yes, CE certification/FDA approval with ISO/IEC documentation would be considered as proof of regulatory evidence.
4.	Can local technical support and aftersales obligations be fulfilled through an authorised Ugandan sub-contractor/service partner?	Yes, technical support and aftersales services may be undertaken by authorised Ugandan sub-contractor/service partner. However, the technical expert should meet

		the minimum qualifications and experience (proof of training by the equipment manufacturer of the proposed model). Please see Section 14, “Qualitative Selection” of the tender document. Additionally, if a local company is nominated by the tenderer, proof of a valid/existing agreement should be provided.
--	--	---